



भारतीय आयुर्विज्ञान अनुसंधान परिषद
स्वास्थ्य अनुसंधान विभाग, स्वास्थ्य और परिवार
कल्याण मंत्रालय, भारत सरकार

Indian Council of Medical Research
Department of Health Research, Ministry of Health
and Family Welfare, Government of India

Date: 29.01.2022

ICMR APPROVED REAL TIME PCR (RT-PCR) ASSAYS FOR DETECTION OF OMICRON VARIANT OF SARS-COV-2

S. No	Name of company	Name of the kit	*Lot no. / Batch No.
1.	TATA Medical & Diagnostics Ltd., Mumbai (Maharashtra), India	TATA MD CHECK RT-PCR OmiSure	K035-V001
2.	Kriya Medical Technologies Pvt. Ltd.	KIRIDA Novus SARS-CoV-2 Real-Time PCR kit- Omicron detection kit	OD060122/01

- Due to the rapidly evolving nature of SARS CoV-2 Omicron variant, sensitivity and specificity of the above kits may change as compared to that reported at the time of validation.
- All samples testing negative for Omicron by the above kits may require further confirmation by sequencing.
- The above kits have been validated for the mentioned batch number only, as per manufacturer's instructions. Responsibility for batch-to-batch consistency lies with the manufacturer.